







(Full quality assurance system)

This is to certify that the company

PIXMEO SARL

266 rue de Bernex 1233 Bernex Switzerland

has implemented and maintains a full quality assurance system which applies to the products at every stage from design to final controls.

Through an audit, documented in a report, performed by DQS Medizinprodukte GmbH, it was verified that the management system fulfills the requirements of

Annex II – excluding Section 4 of Council Directive 93/42/EEC concerning medical devices

with respect to the following medical devices:

The manufacturer is subject to surveillance according to Annex II, Section 5. The CE marking with the Notified Body Identification Number (0297) may be affixed on the devices listed in the certificate. An EC Design Examination Certificate according to Annex II, Section 4 is required for class III devices covered by this certificate. The certificate is in the case of class I(s) devices (I(s) = class I products placed on the market in sterile conditions) limited to the aspects of manufacture concerned with securing and maintaining sterile conditions. The certificate is in the case of class I(m) devices (I(m) = class I devices with a measuring function) limited to the aspects of manufacture concerned with the conformity of the products with the metrological requirements.

Certificate registration no. 527201 MR2
Certificate unique ID 170649404
Effective date 2017-01-23
Expiry date 2022-01-22
Frankfurt am Main 2017-01-23

DQS Medizinprodukte GmbH

Molen

Sigrid Uhlemann Managing Director Dr. Thomas Feldmann Head of Certification Body

August-Schanz-Straße 21, 60433 Frankfurt am Main, Tel. +49 (0) 69 95427-300, medical.devices@dqs-med.de



20160501/3/EN380308